



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/785,304	02/24/2004	Dean Vinson Davis	2004P00817US01	2103

7590 03/17/2009
Siemens Corporation
Intellectual Property Department
170 Wood Avenue South
Iselin, NJ 08830

EXAMINER

SIEFKE, SAMUEL P

ART UNIT	PAPER NUMBER
----------	--------------

1797

MAIL DATE	DELIVERY MODE
-----------	---------------

03/17/2009

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No.	Applicant(s)	
	10/785,304	DAVIS ET AL.	
	Examiner	Art Unit	
	SAM P. SIEFKE	1797	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 3/5/09.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-4 and 7-10 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-4, 7-8, and 10 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Objections

Claim 9 is objected to because of the following informalities: The claim reads "includes second vacuum pump" when the claim should read "includes a second vacuum pump". Appropriate correction is required.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 1-4, 7-8 and 10 are rejected under 35 U.S.C. 102(e) as being anticipated by Taylor et al. (USPN 6,670,608).

Taylor teaches a gas sampling method for a mass spectrometer that comprises evacuating low temperature zone (not heated by heater 44, so inherently lower than the zone in which heater 44 heats) of a sampling system using a vacuum pump (mass spectrometer operating 10 to the minus 6 torr, fig. 3) (col. 3, lines 5-10), admitting a portion of a high temperature (heater 44 heats portion of capillary 32) process stream

Art Unit: 1797

into the mass spectrometer through an orifice (the point in which the capillary 32 passes through wall of the vacuum chamber), maintaining a stable vacuum pressure in the mass spectrometer (inherent for proper functioning of mass spectrometer), introducing a sample from the low temperature zone of the sampling system into a mass spectrometer through a sample introduction valve (the actual introduction through the valve of the mass spectrometer), maintaining a temperature of the low temperature zone above a boiling point of the target sample component at the stable vacuum pressure. It is inherent that the sample is maintained above a boiling point (gaseous state) during analysis in the mass spectrometer because this is how mass spectrometer provides analysis.

Allowable Subject Matter

Claim 9 is objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Response to Arguments

Applicant's arguments filed 3/5/09 have been fully considered but they are not persuasive. Applicant argues, " Specifically, claim 1, from which each of claims 2-4, 7-9, and 10 depends, state, inter alia, yet no substantial evidence has been presented that the applied portions of the cited relied- upon references teach, "the test equipment comprising a mass spectrometer" or "evacuating with a second vacuum pump a

Art Unit: 1797

chamber of the mass spectrometer to a pressure lower than the stable vacuum pressure in the low temperature zone"." Regarding the argument that the prior art does not teach the test equipment comprising a mass spectrometer. The Examiner points to figure 3 and 5 which specifically show a vacuum chamber and a mass spectrometer. Further, the title specifically states "Gas sampling for mass spectrometer" and as seen in column 3, lines 28-34. Further claim 1 does not even require a mass spectrometer as claimed.

Regarding the argument that the prior art does not disclose "evacuating with a second vacuum pump a chamber of the mass spectrometer to a pressure lower than the stable vacuum pressure in the low temperature zone", the Examiner states that this limitation has been deleted from claim 1 in the amendment dated 3/5/09 and this claim limitation only is seen in the indicated allowable subject matter seen in claim 9. Thus the argument is moot.

Conclusion

THIS ACTION IS MADE FINAL even though it is a first action in this case. See MPEP § 706.07(b). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the

Art Unit: 1797

shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no, however, event will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to SAM P. SIEFKE whose telephone number is (571)272-1262. The examiner can normally be reached on M-F 7:00am-5:00pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill A. Warden can be reached on 571-272-1700. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Samuel P Siefke/
Primary Examiner, Art Unit 1797

Application/Control Number: 10/785,304
Art Unit: 1797

Page 6